

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260D

Datafile : VY020199.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1107SBSD01	Dichlorodifluoromethane	20	19.2	ug/Kg	96	3		64	136	20
	Chloromethane	20	19.6	ug/Kg	98	6		70	130	20
	Vinyl chloride	20	20.0	ug/Kg	100	7		72	129	20
	Bromomethane	20	20.8	ug/Kg	104	7		58	141	20
	Chloroethane	20	20.5	ug/Kg	103	10		69	130	20
	Trichlorofluoromethane	20	19.6	ug/Kg	98	12		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	19.9	ug/Kg	100	8		81	123	20
	1,1-Dichloroethene	20	19.1	ug/Kg	96	10		79	121	20
	Acetone	100	110	ug/Kg	110	10		60	131	20
	Carbon disulfide	20	17.4	ug/Kg	87	12		45	154	20
	Methyl tert-butyl Ether	20	21.8	ug/Kg	109	1		77	129	20
	Methyl Acetate	20	22.8	ug/Kg	114	2		69	149	20
	Methylene Chloride	20	19.2	ug/Kg	96	6		56	174	20
	trans-1,2-Dichloroethene	20	19.1	ug/Kg	96	9		80	123	20
	1,1-Dichloroethane	20	20.7	ug/Kg	104	2		82	123	20
	Cyclohexane	20	17.9	ug/Kg	90	11		76	122	20
	2-Butanone	100	110	ug/Kg	110	0		69	131	20
	Carbon Tetrachloride	20	19.0	ug/Kg	95	11		76	129	20
	cis-1,2-Dichloroethene	20	19.9	ug/Kg	100	5		82	123	20
	Bromochloromethane	20	19.9	ug/Kg	100	4		80	127	20
	Chloroform	20	20.6	ug/Kg	103	5		82	125	20
	1,1,1-Trichloroethane	20	20.3	ug/Kg	102	5		80	126	20
	Methylcyclohexane	20	17.5	ug/Kg	88	13		77	123	20
	Benzene	20	19.9	ug/Kg	100	5		84	121	20
	1,2-Dichloroethane	20	21.1	ug/Kg	106	4		81	126	20
	Trichloroethene	20	19.7	ug/Kg	99	5		83	122	20
	1,2-Dichloropropane	20	20.8	ug/Kg	104	5		83	122	20
	Bromodichloromethane	20	20.6	ug/Kg	103	1		82	123	20
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	0		70	135	20
	Toluene	20	19.9	ug/Kg	100	7		83	122	20
	t-1,3-Dichloropropene	20	19.8	ug/Kg	99	5		78	124	20
	cis-1,3-Dichloropropene	20	20.3	ug/Kg	102	4		81	122	20
	1,1,2-Trichloroethane	20	21.5	ug/Kg	108	6		82	125	20
	2-Hexanone	100	110	ug/Kg	110	0		66	138	20
	Dibromochloromethane	20	21.4	ug/Kg	107	1		79	125	20
	1,2-Dibromoethane	20	21.1	ug/Kg	106	1		80	125	20
	Tetrachloroethene	20	19.7	ug/Kg	99	6		83	125	20
	Chlorobenzene	20	20.1	ug/Kg	101	4		84	122	20
	Ethyl Benzene	20	19.4	ug/Kg	97	7		82	124	20
	m/p-Xylenes	40	39.0	ug/Kg	98	7		83	124	20
	o-Xylene	20	19.4	ug/Kg	97	9		83	123	20
	Styrene	20	19.9	ug/Kg	100	7		82	124	20
	Bromoform	20	21.9	ug/Kg	110	1		75	127	20
	Isopropylbenzene	20	19.4	ug/Kg	97	9		82	124	20
	1,1,2,2-Tetrachloroethane	20	21.1	ug/Kg	106	0		77	127	20
	1,3-Dichlorobenzene	20	20.0	ug/Kg	100	8		83	122	20
	1,4-Dichlorobenzene	20	20.0	ug/Kg	100	8		84	121	20
	1,2-Dichlorobenzene	20	20.5	ug/Kg	103	6		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4722

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260D

Datafile : VY020199.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY1107SBSD01	1,2-Dibromo-3-Chloropropane	20	20.6	ug/Kg	103	4		66	134	20
	1,2,4-Trichlorobenzene	20	19.5	ug/Kg	98	6		78	127	20
	1,2,3-Trichlorobenzene	20	19.5	ug/Kg	98	2		70	137	20