

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WALS01
 Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722
 Lab File ID: VY020196.D Date Analyzed: 11/07/2024
 Instrument ID: MSVOA_Y Time Analyzed: 10:27
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	221675	7.71	394769	8.62	345161	11.41
UPPER LIMIT	443350	8.213	789538	9.116	690322	11.914
LOWER LIMIT	110838	7.213	197385	8.116	172581	10.914
EPA SAMPLE NO.						
WC-2 (2)	198592	7.71	428344	8.62	407022	11.41
WC-3 (5)	142540	7.71	298274	8.62	289563	11.41
WC-1 (3.5)	181277	7.71	389130	8.62	362163	11.41
WC-2 (4)	179581	7.71	387425	8.62	366025	11.41
WC-3 (3)	175489	7.71	372455	8.62	352192	11.41
VY1107SBL01	168265	7.71	343319	8.62	313025	11.41
VY1107SBS01	202685	7.71	361864	8.62	315230	11.41
VY1107SBSD01	206563	7.71	375441	8.62	325231	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.