

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WALS01
Lab Code: CHEM Case No.: P4722 SAS No.: P4722 SDG NO.: P4722
Lab File ID: VY020221.D Date Analyzed: 11/08/2024
Instrument ID: MSVOA_Y Time Analyzed: 09:33
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	282936	7.71	514187	8.62	441792	11.42
UPPER LIMIT	565872	8.213	1028370	9.116	883584	11.92
LOWER LIMIT	141468	7.213	257094	8.116	220896	10.92
EPA SAMPLE NO.						
WC-1(5.5)	209744	7.71	444878	8.62	397795	11.41
VY1108SBL01	212436	7.71	452508	8.62	403401	11.42
VY1108SBS01	263593	7.71	480169	8.62	412478	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.