

Laboratory Control Sample Summary

Client: Tetra Tech, EMI

SDG No.: P4723

Project: R36720

Run No.: LB133384

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133384BS							
TSS	mg/L	550	536		98	1	90-110	11/07/2024