

Pratik Patel







From: Sohil Jodhani
Sent: Monday, November 25, 2024 1:55 PM
To: Panday, Prakriti P
Cc: Nimisha Pandya; Mohammad Ahmed
Subject: RE: REGION 2 | CASE 51825 | LAB ACE | SDG BH9C6 & BH9B8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Hi Prakriti,

Yes, All analysis are within holding time.

Thanks & Regards,



Sohil Jodhani
QA/QC Director
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3152
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com      

From: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Sent: Monday, November 25, 2024 1:44 PM
To: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>
Subject: REGION 2 | CASE 51825 | LAB ACE | SDG BH9C6 & BH9B8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,

Please see the Region's follow up question below regarding holding time for Case 51825/SDGs BH9C6 & BH9B.

ISSUE: Lab has received soil samples for VOA analysis. Lab has analyzed undiluted VOA analysis for the samples BH9G6, BH9D2, BH9E4 & BH9E3. Samples had internal standard recoveries outside the QC limits as you can see attached internal standard summary for your reference. As a corrective action, Lab has re-analyzed these samples and there was some technical error in instrument and closing CCV was not injected for the analytical sequence. Therefore, as a further corrective action, Lab has performed the medium level VOA analysis for these samples. In this case, Lab would like to confirm that lab will report all analysis for final electronic deliverables.

Please see attachment for your reference.

REGION: It is OK with the region to have all data. Please ensure SOW requirement. Are all analyses within holding time ?
Thanks.

Thank you.

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261
GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: December 4th and 5th

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Kumar, Narendra <Kumar.Narendra@epa.gov>
Sent: Friday, November 22, 2024 11:08 AM
To: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Cc: Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Arnone, Russell <Arnone.Russell@epa.gov>
Subject: RE: REGION 2 | CASE 51825 | LAB ACE | SDG BH9C6 & BH9B8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello Prakriti,
It is OK with the region to have all data. Please ensure SOW requirement. Are all analyses within holding time ?
Thanks.

Narendra Kumar

From: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Sent: Friday, November 22, 2024 9:04 AM
To: Kumar, Narendra <Kumar.Narendra@epa.gov>; Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Arnone, Russell <Arnone.Russell@epa.gov>
Subject: REGION 2 | CASE 51825 | LAB ACE | SDG BH9C6 & BH9B8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning,

The laboratory is awaiting the Region's response for Case 51825/ SDGs BH9C6 & BH9B8.

Thank you,

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261
GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



Leave Alert: December 4th and 5th

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Panday, Prakriti P
Sent: Tuesday, November 19, 2024 4:07 PM
To: REGION_2 (Kumar.Narendra@epa.gov) <Kumar.Narendra@epa.gov>; Leung, Christina <leung.christina@epa.gov>; Region_2 (Feranda.Jennifer@epa.gov) <Feranda.Jennifer@epa.gov>; Region_2 (Brandon-Bazile.Kim@epa.gov) <Brandon-Bazile.Kim@epa.gov>; REGION_2 (Arnone.Russell@epa.gov) <Arnone.Russell@epa.gov>
Subject: REGION 2 | CASE 51825 | LAB ACE | SDG BH9C6 & BH9B8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Good afternoon,

Regarding Case 51825/SDGs BH9C6 & BH9B8, please advise on the laboratory proposal below.

ISSUE: Lab has received soil samples for VOA analysis. Lab has analyzed undiluted VOA analysis for the samples BH9G6, BH9D2, BH9E4 & BH9E3. Samples had internal standard recoveries outside the QC limits as you can see attached internal standard summary for your reference. As a corrective action, Lab has re-analyzed these samples and there was some technical error in instrument and closing CCV was not injected for the analytical sequence. Therefore, as a further corrective action, Lab has performed the medium level VOA analysis for these samples. In this case, Lab would like to confirm that lab will report all analysis for final electronic deliverables.

Please see attachment for your reference.

Thank you,

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261
GDIT Federal Civilian Division
Under Contract to EPA

15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



Leave Alert: December 4th and 5th

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASsupport

Sent: Tuesday, November 19, 2024 3:56 PM

To: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: REGION 2 | CASE 51825 | LAB ACE | SDG BH9C6 & BH9B8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team

DASsupport@gdit.com

GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>

Sent: Tuesday, November 19, 2024 1:14 PM

To: DASsupport <DASsupport@gdit.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: REGION 2 | CASE 51825 | LAB ACE | SDG BH9C6 & BH9B8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello,

Lab is sending this email with regards to case 51825.

Lab has received soil samples for VOA analysis. Lab has analyzed undiluted VOA analysis for the samples BH9G6, BH9D2, BH9E4 & BH9E3. Samples had internal standard recoveries outside the QC limits as you can see attached internal standard summary for your reference. As a corrective action, Lab has re-analyzed these samples and there was some technical error in instrument and closing CCV was not injected for the analytical sequence. Therefore, as a further corrective action, Lab has performed the medium level VOA analysis for these samples. In this case, Lab would like to confirm that lab will report all analysis for final electronic deliverables.

Please see attachment for your reference.

Thanks & Regards,



Sohil Jodhani
QA/QC Director
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3152
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

