

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51850 MA No.: SDG No.: MC0B82
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
		Cr	Cu	Mn	Mo	Ni
Calcium	373.69	0	0	0	0	0
Magnesium	279.079	0	0	0	0	0

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51850 MA No.: SDG No.: MC0B82
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
		Pb	Ti	V	Zn	
Calcium	373.69	0	0	0	0	
Magnesium	279.079	0	0	0	0	

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51850 MA No.: SDG No.: MC0B82
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
Calcium	373.69	0	0	0	0	
Magnesium	279.079	0	0	0	0	