

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51850 MA No.: _____ SDG No.: MC0B82
Instrument ID: P5 Start Date: 11/12/2024
Analytical Method: ICP-AES Run Batch: LB133417
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Calcium	0	-0.61	0	5000	5500	10	62500	62000	-1
Magnesium	0	0.15	0	5000	5200	4	62500	58000	-7

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51850 MA No.: SDG No.: MC0B82
Instrument ID: P5 Start Date: 11/12/2024
Analytical Method: ICP-AES Run Batch: LB133417
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Calcium	125000	130000	4	250000	260000	4	500000	510000	2
Magnesium	125000	120000	-4	250000	250000	0	500000	510000	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51850 MA No.: SDG No.: MC0B82
Instrument ID: P5 Start Date: 11/12/2024
Analytical Method: ICP-AES Run Batch: LB133417
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Calcium	800000	780000	-3						
Magnesium	800000	790000	-1						