

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51850 MA No. : SDG No.: MC0B82
 Initial Calibration Verification Source : MP83143
 Continuing Calibration Verification Source : MP83146
 Run Batch: LB133417 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV010				ID: CCV064				ID: CCV065		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Calcium	10000	10000	102	0	400000	420000	104	0	410000	102	0
Magnesium	6000	5500	91	2	400000	410000	103	0	410000	102	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51850 MA No. : SDG No.: MC0B82
 Initial Calibration Verification Source : MP83143
 Continuing Calibration Verification Source : MP83146
 Run Batch: LB133417 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV066				ID: CCV067		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Calcium					400000	410000	101	0	410000	103	1
Magnesium					400000	400000	101	0	410000	102	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51850 MA No. : SDG No.: MC0B82
 Initial Calibration Verification Source : MP83143
 Continuing Calibration Verification Source : MP83146
 Run Batch: LB133417 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV068				ID: CCV069		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Calcium					400000	410000	104	0	420000	105	0
Magnesium					400000	410000	103	0	410000	103	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51850 MA No. : SDG No.: MC0B82
 Initial Calibration Verification Source : MP83143
 Continuing Calibration Verification Source : MP83146
 Run Batch: LB133417 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV070				ID: CCV071		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Calcium					400000	420000	105	0	430000	107	0
Magnesium					400000	420000	104	0	420000	105	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51850 MA No. : SDG No.: MC0B82
 Initial Calibration Verification Source : MP83143
 Continuing Calibration Verification Source : MP83146
 Run Batch: LB133417 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV072				ID: CCV073		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Calcium					400000	420000	104	0	420000	105	0
Magnesium					400000	410000	103	0	410000	104	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51850 MA No. : SDG No.: MC0B82
 Initial Calibration Verification Source : MP83143
 Continuing Calibration Verification Source : MP83146
 Run Batch: LB133417 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV074				ID: CCV075		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Calcium					400000	430000	106	0	420000	106	1
Magnesium					400000	420000	104	0	420000	104	1