

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51850 MA No.: SDG No.: MC0B82
Instrument ID: CV1 Start Date: 11/10/2024
Analytical Method: CVAA Run Batch: LB133376
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0070	0	0.2	0.17	-17	2.5	2.4	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51850 MA No.: _____ SDG No.: MC0B82
Instrument ID: CV1 Start Date: 11/10/2024
Analytical Method: CVAA Run Batch: LB133376
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.1	2	7.5	7.6	2	10	9.9	-1