EPA SAMPLE NO.

MC0B86S

## FORM 5A - IN MATRIX SPIKE SAMPLE RECOVERY

Lab Name:	Chemtech Con	nsulting Group	Contract:	EPW14030			
Lab Code:	ACE	Case No.: <u>51850</u>	MA No. :		SDG No.:	MC0B82	
Matrix :	Water		Analytical	Method:	CVAA		
% Solids:							
Concentrat	ion Units (μg	/L, mg/L or mg/kg dry	weight): ug/I				

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	%R	Q
Mercury	75 - 125	0.83		0.2	U	1.0	83	