

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Furino and Sons, Inc. **SDG No.:** P4732
Contract: FURI01 **Lab Code:** CHEM **Case No.:** P4732 **SAS No.:** P4732

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164757BS							
Aluminum	mg/Kg	97.1	96.3		99	80 - 120	P
Antimony	mg/Kg	38.8	37.8		97	80 - 120	P
Arsenic	mg/Kg	38.8	38.2		98	80 - 120	P
Barium	mg/Kg	9.7	8.93		92	80 - 120	P
Beryllium	mg/Kg	9.7	9.44		97	80 - 120	P
Cadmium	mg/Kg	9.7	8.83		91	80 - 120	P
Calcium	mg/Kg	48.5	47.6	J	98	80 - 120	P
Chromium	mg/Kg	19.4	17.6		91	80 - 120	P
Cobalt	mg/Kg	9.7	9.14		94	80 - 120	P
Copper	mg/Kg	14.6	14.8		101	80 - 120	P
Iron	mg/Kg	150	131		87	80 - 120	P
Lead	mg/Kg	48.5	45.7		94	80 - 120	P
Magnesium	mg/Kg	97.1	91.6	J	94	80 - 120	P
Manganese	mg/Kg	9.7	9.16		94	80 - 120	P
Nickel	mg/Kg	24.3	22.9		94	80 - 120	P
Potassium	mg/Kg	490	448		91	80 - 120	P
Selenium	mg/Kg	97.1	95.4		98	80 - 120	P
Silver	mg/Kg	3.6	3.54		98	80 - 120	P
Sodium	mg/Kg	150	120		80	80 - 120	P
Thallium	mg/Kg	97.1	97.3		100	80 - 120	P
Vanadium	mg/Kg	14.6	14.0		96	80 - 120	P
Zinc	mg/Kg	9.7	9.96		103	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Furino and Sons, Inc.

SDG No.: P4732

Contract: FURI01

Lab Code: CHEM

Case No.: P4732

SAS No.: P4732

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164762BS Mercury	mg/Kg	0.26	0.29		110	80 - 120	CV