

CALIBRATION VERIFICATION SUMMARY

Contract: FURI01

Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG NO.: P4732

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107769.D Time Analyzed: 09:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.423	5.623	533.240	500.000	6.6
Aroclor-1016-2	5.547	5.445	5.645	524.920	500.000	5.0
Aroclor-1016-3	5.610	5.507	5.707	512.270	500.000	2.5
Aroclor-1016-4	5.707	5.604	5.804	536.660	500.000	7.3
Aroclor-1016-5	6.003	5.899	6.099	549.900	500.000	10.0
Aroclor-1260-1	7.133	7.026	7.226	569.760	500.000	14.0
Aroclor-1260-2	7.389	7.282	7.482	591.260	500.000	18.3
Aroclor-1260-3	7.752	7.643	7.843	572.350	500.000	14.5
Aroclor-1260-4	7.978	7.868	8.068	583.920	500.000	16.8
Aroclor-1260-5	8.293	8.181	8.381	596.880	500.000	19.4
Decachlorobiphenyl	10.079	9.958	10.158	51.570	50.000	3.1
Tetrachloro-m-xylene	4.373	4.274	4.474	54.530	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Contract: FURI01

Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG NO.: P4732

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107769.D Time Analyzed: 09:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.626	4.826	561.030	500.000	12.2
Aroclor-1016-2	4.742	4.645	4.845	577.570	500.000	15.5
Aroclor-1016-3	4.919	4.821	5.021	552.200	500.000	10.4
Aroclor-1016-4	4.960	4.862	5.062	485.110	500.000	-3.0
Aroclor-1016-5	5.173	5.076	5.276	550.490	500.000	10.1
Aroclor-1260-1	6.203	6.107	6.307	559.730	500.000	11.9
Aroclor-1260-2	6.391	6.294	6.494	577.550	500.000	15.5
Aroclor-1260-3	6.545	6.448	6.648	543.970	500.000	8.8
Aroclor-1260-4	7.016	6.918	7.118	540.980	500.000	8.2
Aroclor-1260-5	7.257	7.159	7.359	574.450	500.000	14.9
Decachlorobiphenyl	8.638	8.539	8.739	56.920	50.000	13.8
Tetrachloro-m-xylene	3.643	3.545	3.745	57.040	50.000	14.1