

CALIBRATION VERIFICATION SUMMARY

Contract: FURI01

Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG NO.: P4732

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107797.D Time Analyzed: 21:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.524	5.423	5.623	561.890	500.000	12.4
Aroclor-1016-2	5.546	5.445	5.645	544.590	500.000	8.9
Aroclor-1016-3	5.609	5.507	5.707	539.140	500.000	7.8
Aroclor-1016-4	5.706	5.604	5.804	559.970	500.000	12.0
Aroclor-1016-5	6.001	5.899	6.099	550.930	500.000	10.2
Aroclor-1260-1	7.131	7.026	7.226	506.260	500.000	1.3
Aroclor-1260-2	7.388	7.282	7.482	536.110	500.000	7.2
Aroclor-1260-3	7.751	7.643	7.843	551.140	500.000	10.2
Aroclor-1260-4	7.975	7.868	8.068	569.480	500.000	13.9
Aroclor-1260-5	8.290	8.181	8.381	568.620	500.000	13.7
Decachlorobiphenyl	10.077	9.958	10.158	48.480	50.000	-3.0
Tetrachloro-m-xylene	4.372	4.274	4.474	56.680	50.000	13.4

CALIBRATION VERIFICATION SUMMARY

Contract: FURI01

Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG NO.: P4732

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107797.D Time Analyzed: 21:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.626	4.826	588.760	500.000	17.8
Aroclor-1016-2	4.742	4.645	4.845	598.940	500.000	19.8
Aroclor-1016-3	4.918	4.821	5.021	576.400	500.000	15.3
Aroclor-1016-4	4.960	4.862	5.062	516.750	500.000	3.4
Aroclor-1016-5	5.173	5.076	5.276	571.700	500.000	14.3
Aroclor-1260-1	6.204	6.107	6.307	555.680	500.000	11.1
Aroclor-1260-2	6.391	6.294	6.494	584.000	500.000	16.8
Aroclor-1260-3	6.545	6.448	6.648	554.980	500.000	11.0
Aroclor-1260-4	7.015	6.918	7.118	541.250	500.000	8.3
Aroclor-1260-5	7.256	7.159	7.359	566.720	500.000	13.3
Decachlorobiphenyl	8.638	8.539	8.739	50.910	50.000	1.8
Tetrachloro-m-xylene	3.643	3.545	3.745	59.240	50.000	18.5