



Contract:	FURI01					
Lab Code:	CHEM	Case No.:	P4732	SAS No.:	P4732	SDG NO.: P4732
Instrument ID:	ECD_S	Calibration Date(s):		11/08/2024	11/08/2024	
		Calibration Times:		15:13	16:58	

LAB FILE ID:	RT 200 =	<u>PS028356.D</u>	RT 500 =	<u>PS028357.D</u>
RT 750 =	<u>PS028358.D</u>	RT 1000 =	<u>PS028359.D</u>	RT 1500 =
			<u>PS028360.D</u>	

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: FURI01

Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG NO.: P4732

Instrument ID: ECD_S Calibration Date(s): 11/08/2024 11/08/2024

Calibration Times: 15:13 16:58

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS028356.D</u>	RT 500 = <u>PS028357.D</u>
	RT 750 = <u>PS028358.D</u>	RT 1000 = <u>PS028359.D</u>
		RT 1500 = <u>PS028360.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM TO	
2,4,5-TP(Silvex)	9.91	9.91	9.91	9.91	9.91	9.91	9.81	10.01
2,4-D	9.01	9.00	9.01	9.01	9.01	9.00	8.90	9.10
2,4-DCAA	7.76	7.75	7.76	7.76	7.76	7.75	7.65	7.85