

CALIBRATION VERIFICATION SUMMARY

Contract: FURI01

Lab Code: CHEM **Case No.:** P4732 **SAS No.:** P4732 **SDG NO.:** P4732

Continuing Calib Date: 11/13/2024 **Initial Calibration Date(s):** 11/08/2024 11/08/2024

Continuing Calib Time: 04:14 **Initial Calibration Time(s):** 15:13 16:58

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.24	7.25	7.15	7.35	0.01
2,4-D	8.37	8.38	8.28	8.48	0.01
2,4,5-TP(Silvex)	9.25	9.26	9.16	9.36	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: FURI01

Lab Code: CHEM **Case No.:** P4732 **SAS No.:** P4732 **SDG NO.:** P4732

Continuing Calib Date: 11/13/2024 **Initial Calibration Date(s):** 11/08/2024 11/08/2024

Continuing Calib Time: 04:14 **Initial Calibration Time(s):** 15:13 16:58

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.74	7.76	7.66	7.86	0.02
2,4-D	8.99	9.01	8.91	9.11	0.02
2,4,5-TP(Silvex)	9.90	9.91	9.81	10.01	0.01