

CALIBRATION VERIFICATION SUMMARY

Contract: FURI01

Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG NO.: P4732

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL03 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028473.D Time Analyzed: 22:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.250	9.160	9.360	733.770	712.500	3.0
2,4-D	8.371	8.279	8.479	706.280	705.000	0.2
2,4-DCAA	7.241	7.150	7.350	747.510	750.000	-0.3

CALIBRATION VERIFICATION SUMMARY

Contract: FURI01

Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG NO.: P4732

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL03 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028473.D Time Analyzed: 22:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.900	9.811	10.011	667.590	712.500	-6.3
2,4-D	8.993	8.905	9.105	653.200	705.000	-7.3
2,4-DCAA	7.745	7.655	7.855	681.740	750.000	-9.1