

Surrogate Summary

SDG No.: **P4732**

Client: **Furino and Sons, Inc.**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL092652.D	PIBLK-PL092652.D	Decachlorobiphenyl	1	20	22.7	114		43	140
		Tetrachloro-m-xylene	1	20	21.6	108		77	126
		Decachlorobiphenyl	2	20	21.7	109		43	140
		Tetrachloro-m-xylene	2	20	20.4	102		77	126
I.BLK-PL092941.D	PIBLK-PL092941.D	Decachlorobiphenyl	1	20	20.8	104		43	140
		Tetrachloro-m-xylene	1	20	22.3	112		77	126
		Decachlorobiphenyl	2	20	19.8	99		43	140
		Tetrachloro-m-xylene	2	20	21.6	108		77	126
PB164849BL	PB164849BL	Decachlorobiphenyl	1	20	21.1	105		43	140
		Tetrachloro-m-xylene	1	20	20.7	104		77	126
		Decachlorobiphenyl	2	20	20.5	103		43	140
		Tetrachloro-m-xylene	2	20	20.1	101		77	126
PB164849BS	PB164849BS	Decachlorobiphenyl	1	20	19.3	96		43	140
		Tetrachloro-m-xylene	1	20	19.1	95		77	126
		Decachlorobiphenyl	2	20	18.9	94		43	140
		Tetrachloro-m-xylene	2	20	17.9	90		77	126
PB164694TB	PB164694TB	Decachlorobiphenyl	1	20	21.2	106		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	20.8	104		43	140
		Tetrachloro-m-xylene	2	20	20.1	101		77	126
I.BLK-PL092959.D	PIBLK-PL092959.D	Decachlorobiphenyl	1	20	22.0	110		43	140
		Tetrachloro-m-xylene	1	20	22.1	110		77	126
		Decachlorobiphenyl	2	20	22.2	111		43	140
		Tetrachloro-m-xylene	2	20	21.5	107		77	126
I.BLK-PL092971.D	PIBLK-PL092971.D	Decachlorobiphenyl	1	20	21.6	108		43	140
		Tetrachloro-m-xylene	1	20	22.7	114		77	126
		Decachlorobiphenyl	2	20	20.6	103		43	140
		Tetrachloro-m-xylene	2	20	21.6	108		77	126
P4718-03MS	WB-307-SB02MS	Decachlorobiphenyl	1	20	14.1	71		43	140
		Tetrachloro-m-xylene	1	20	24.6	123		77	126
		Decachlorobiphenyl	2	20	13.1	66		43	140
		Tetrachloro-m-xylene	2	20	19.3	96		77	126
P4718-03MSD	WB-307-SB02MSD	Decachlorobiphenyl	1	20	14.4	72		43	140
		Tetrachloro-m-xylene	1	20	24.6	123		77	126
		Decachlorobiphenyl	2	20	13.2	66		43	140
		Tetrachloro-m-xylene	2	20	19.3	97		77	126
P4732-02	PPE-COMP	Decachlorobiphenyl	1	20	20.2	101		43	140
		Tetrachloro-m-xylene	1	20	20.2	101		77	126
		Decachlorobiphenyl	2	20	19.9	99		43	140
		Tetrachloro-m-xylene	2	20	18.4	92		77	126
I.BLK-PL092992.D	PIBLK-PL092992.D	Decachlorobiphenyl	1	20	22.2	111		43	140

Surrogate Summary

SDG No.: P4732

Client: Furino and Sons, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL092992.D	PIBLK-PL092992.D	Tetrachloro-m-xylene	1	20	22.8	114		77	126
		Decachlorobiphenyl	2	20	21.8	109		43	140
		Tetrachloro-m-xylene	2	20	22.2	111		77	126