

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: FURI01

Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG No.: P4732

Instrument ID: BNA_F Calibration Date/Time: 11/08/2024 09:35

Lab File ID: BF140286.D Init. Calib. Date(s): 11/05/2024 11/05/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:26 15:30

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.354	1.280		-5.5	
2-Fluorophenol	1.207	1.151		-4.6	
Phenol-d6	1.552	1.455		-6.3	
1,4-Dichlorobenzene	1.492	1.440		-3.5	20.0
2-Methylphenol	1.053	1.025		-2.7	
3+4-Methylphenols	1.323	1.267		-4.2	
Nitrobenzene-d5	0.402	0.378		-6.0	
Hexachloroethane	0.556	0.530		-4.7	
Nitrobenzene	0.422	0.398		-5.7	
Hexachlorobutadiene	0.229	0.225		-1.7	20.0
2,4,6-Trichlorophenol	0.363	0.362		-0.3	20.0
2-Fluorobiphenyl	1.253	1.125		-10.1	
2,4,5-Trichlorophenol	0.381	0.388		1.8	
2,4-Dinitrotoluene	0.376	0.395		5.1	
2,4,6-Tribromophenol	0.198	0.223		12.6	
Hexachlorobenzene	0.245	0.245		0.0	
Pentachlorophenol	0.132	0.154		16.7	20.0
Terphenyl-d14	1.221	1.161		-4.9	

All other compounds must meet a minimum RRF of 0.010.