

Report of Analysis

Client:	Furino and Sons, Inc.	Date Collected:	11/06/24
Project:	PPE Contamination	Date Received:	11/06/24
Client Sample ID:	TP-14MS	SDG No.:	P4732
Lab Sample ID:	P4739-04MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101557.D	1	11/07/24 11:30	11/08/24 18:07	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	340		15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	520		8.40	50.0	ug/L
95-48-7	2-Methylphenol	560		11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	540		11.5	100	ug/L
67-72-1	Hexachloroethane	520		10.1	50.0	ug/L
98-95-3	Nitrobenzene	540		12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	510		12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	590		8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	580		10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	550		15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	550		11.4	50.0	ug/L
87-86-5	Pentachlorophenol	1300	E	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	175		10 - 139	117%	SPK: 150
13127-88-3	Phenol-d6	155		10 - 134	103%	SPK: 150
4165-60-0	Nitrobenzene-d5	112		49 - 133	112%	SPK: 100
321-60-8	2-Fluorobiphenyl	110		52 - 132	110%	SPK: 100
118-79-6	2,4,6-Tribromophenol	166		44 - 137	110%	SPK: 150
1718-51-0	Terphenyl-d14	115		48 - 125	115%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	47500		7.558		
1146-65-2	Naphthalene-d8	202000		10.326		
15067-26-2	Acenaphthene-d10	131000		14.168		
1517-22-2	Phenanthrene-d10	298000		16.906		
1719-03-5	Chrysene-d12	370000		21.072		
1520-96-3	Perylene-d12	477000		23.351		

Report of Analysis

Client:	Furino and Sons, Inc.	Date Collected:	11/06/24
Project:	PPE Contamination	Date Received:	11/06/24
Client Sample ID:	TP-14MS	SDG No.:	P4732
Lab Sample ID:	P4739-04MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BE101557.D	1	11/07/24 11:30	11/08/24 18:07	PB164765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products