

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG NO.: P4732

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/08/2024

Lab File ID: BE101544.D Time Analyzed: 10:07

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	51174	7.558	237509	10.33	179651	14.17
UPPER LIMIT	102348	8.058	475018	10.826	359302	14.674
LOWER LIMIT	25587	7.058	118755	9.826	89825.5	13.674
EPA SAMPLE NO.						
01 PB164765BS	39669	7.56	182397	10.33	121151	14.17
02 PB164694TB	34932	7.56	149600	10.33	99908	14.17
03 PPE-COMP	48513	7.56	202387	10.33	136414	14.17
04 TP-14MS	47543	7.56	202399	10.33	130758	14.17
05 TP-14MSD	43513	7.56	182419	10.33	120731	14.17

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.