

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: FURI01

Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG No.: P4732

Instrument ID: MSVOA\_N Calibration Date/Time: 11/13/2024 09:52

Lab File ID: VN084815.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:46 13:45

GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND              | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|-----------------------|-------|--------|------------|-------|-------|
| Vinyl Chloride        | 0.618 | 0.665  |            | 7.61  | 20    |
| 1,1-Dichloroethene    | 0.569 | 0.539  |            | -5.27 | 20    |
| 2-Butanone            | 0.337 | 0.325  |            | -3.56 | 20    |
| Carbon Tetrachloride  | 0.525 | 0.567  |            | 8     | 20    |
| Chloroform            | 1.121 | 1.116  |            | -0.45 | 20    |
| Benzene               | 1.507 | 1.535  |            | 1.86  | 20    |
| 1,2-Dichloroethane    | 0.488 | 0.515  |            | 5.53  | 20    |
| Trichloroethene       | 0.348 | 0.357  |            | 2.59  | 20    |
| Tetrachloroethene     | 0.333 | 0.370  |            | 11.11 | 20    |
| Chlorobenzene         | 1.119 | 1.142  | 0.3        | 2.06  | 20    |
| 1,2-Dichloroethane-d4 | 0.722 | 0.681  |            | -5.68 | 20    |
| Dibromofluoromethane  | 0.338 | 0.348  |            | 2.96  | 20    |
| Toluene-d8            | 1.247 | 1.247  |            | 0     | 20    |
| 4-Bromofluorobenzene  | 0.466 | 0.482  |            | 3.43  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.