

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4732

Client: Furino and Sons, Inc.

Analytical Method: SW8260-Low

Datafile : VN084822.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1113WBS01	Vinyl chloride	20	20.1	ug/L	101			65	117	
	1,1-Dichloroethene	20	17.5	ug/L	88			74	110	
	2-Butanone	100	91.6	ug/L	92			65	122	
	Carbon Tetrachloride	20	20.1	ug/L	101			77	113	
	Chloroform	20	18.9	ug/L	95			79	113	
	Benzene	20	18.5	ug/L	93			82	109	
	1,2-Dichloroethane	20	19.0	ug/L	95			80	115	
	Trichloroethene	20	19.2	ug/L	96			77	113	
	Tetrachloroethene	20	20.6	ug/L	103			67	123	
	Chlorobenzene	20	18.9	ug/L	95			82	109	