

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4732

Client: Furino and Sons, Inc.

Analytical Method: SW8260-Low

Datafile : VN084823.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1113WBSD01	Vinyl chloride	20	21.9	ug/L	110	9		65	117	20
	1,1-Dichloroethene	20	19.8	ug/L	99	12		74	110	20
	2-Butanone	100	100	ug/L	100	8		65	122	20
	Carbon Tetrachloride	20	21.9	ug/L	110	9		77	113	20
	Chloroform	20	21.5	ug/L	108	13		79	113	20
	Benzene	20	20.7	ug/L	104	11		82	109	20
	1,2-Dichloroethane	20	21.5	ug/L	108	13		80	115	20
	Trichloroethene	20	22.3	ug/L	112	15		77	113	20
	Tetrachloroethene	20	22.0	ug/L	110	7		67	123	20
	Chlorobenzene	20	20.5	ug/L	103	8		82	109	20