

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084571.D
 Acq On : 30 Oct 2024 12:09
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/04/2024
 Supervised By :Mahesh Dadoda 11/04/2024

Quant Time: Oct 31 18:18:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:18:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	186594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	311554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	279228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	138307	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	134521	49.914	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.820%		
35) Dibromofluoromethane	8.165	113	107131	50.801	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.600%		
50) Toluene-d8	10.565	98	405874	52.247	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.500%		
62) 4-Bromofluorobenzene	12.847	95	153622	52.913	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	105.820%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	103092	48.061	ug/l	97
3) Chloromethane	2.365	50	125344	49.788	ug/l	95
4) Vinyl Chloride	2.512	62	112961	48.962	ug/l	95
5) Bromomethane	2.901	94	55194	45.153	ug/l	100
6) Chloroethane	3.071	64	70126	49.078	ug/l	96
7) Trichlorofluoromethane	3.471	101	178947	47.086	ug/l	97
8) Diethyl Ether	3.953	74	65946	49.190	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.353	101	103864	48.366	ug/l	98
10) Methyl Iodide	4.583	142	137274	47.829	ug/l	97
11) Tert butyl alcohol	5.536	59	80598	226.565	ug/l	100
12) 1,1-Dichloroethene	4.324	96	100445	47.270	ug/l	96
13) Acrolein	4.171	56	82129	231.524	ug/l	99
14) Allyl chloride	5.012	41	171045	49.635	ug/l	92
15) Acrylonitrile	5.718	53	252853	245.525	ug/l	98
16) Acetone	4.424	43	189949	243.925	ug/l	99
17) Carbon Disulfide	4.700	76	299041	45.698	ug/l	99
18) Methyl Acetate	5.024	43	139826	49.097	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	328099	50.376	ug/l	97
20) Methylene Chloride	5.271	84	112282	47.370	ug/l	92
21) trans-1,2-Dichloroethene	5.777	96	105050	48.146	ug/l	96
22) Diisopropyl ether	6.671	45	346126	50.549	ug/l	98
23) Vinyl Acetate	6.600	43	1212022	256.673	ug/l	99
24) 1,1-Dichloroethane	6.559	63	198817	48.362	ug/l	100
25) 2-Butanone	7.483	43	294207	233.994	ug/l	98
26) 2,2-Dichloropropane	7.483	77	175511	49.052	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	123436	48.450	ug/l	97
28) Bromochloromethane	7.812	49	95269	58.166	ug/l	95
29) Tetrahydrofuran	7.835	42	206980	248.139	ug/l	97
30) Chloroform	7.965	83	202589	48.417	ug/l	98
31) Cyclohexane	8.253	56	175038	47.043	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	184986	48.573	ug/l	98
36) 1,1-Dichloropropene	8.365	75	140593	48.071	ug/l	99
37) Ethyl Acetate	7.553	43	131332	49.493	ug/l	99
38) Carbon Tetrachloride	8.359	117	160134	48.985	ug/l	97
39) Methylcyclohexane	9.600	83	158489	53.267	ug/l	99
40) Benzene	8.600	78	451074	48.024	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
 Data File : VN084571.D
 Acq On : 30 Oct 2024 12:09
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Quant Time: Oct 31 18:18:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:18:08 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/04/2024
 Supervised By :Mahesh Dadoda 11/04/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	76791	53.814	ug/l	97
42) 1,2-Dichloroethane	8.665	62	153960	50.614	ug/l	100
43) Isopropyl Acetate	8.688	43	235517	48.004	ug/l	99
44) Trichloroethene	9.353	130	104498	48.241	ug/l	98
45) 1,2-Dichloropropane	9.618	63	108302	49.154	ug/l	95
46) Dibromomethane	9.706	93	76019	48.895	ug/l	97
47) Bromodichloromethane	9.888	83	162307	49.525	ug/l	97
48) Methyl methacrylate	9.677	41	107828	52.212	ug/l	98
49) 1,4-Dioxane	9.694	88	37719	1030.669	ug/l #	82
51) 4-Methyl-2-Pentanone	10.447	43	658351	260.254	ug/l	98
52) Toluene	10.629	92	280034	50.978	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	169085	49.843	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	180909	50.386	ug/l	97
55) 1,1,2-Trichloroethane	11.018	97	102401	49.487	ug/l	98
56) Ethyl methacrylate	10.871	69	168060	56.150	ug/l	96
57) 1,3-Dichloropropane	11.159	76	175929	49.539	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	407857	291.210	ug/l	97
59) 2-Hexanone	11.194	43	485391	263.615	ug/l	98
60) Dibromochloromethane	11.359	129	122816	52.078	ug/l	98
61) 1,2-Dibromoethane	11.465	107	103901	49.034	ug/l	98
64) Tetrachloroethene	11.100	164	87457	47.088	ug/l	96
65) Chlorobenzene	11.888	112	296199	47.416	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	104844	48.252	ug/l	99
67) Ethyl Benzene	11.965	91	528107	50.646	ug/l	100
68) m/p-Xylenes	12.070	106	406816	103.053	ug/l	98
69) o-Xylene	12.394	106	192710	52.183	ug/l	99
70) Styrene	12.412	104	336570	52.245	ug/l	99
71) Bromoform	12.576	173	82245	50.303	ug/l #	99
73) Isopropylbenzene	12.694	105	493742	50.942	ug/l	99
74) N-ethyl acetate	12.494	43	211326	51.228	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	148441	45.884	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	122649m	44.939	ug/l	
77) Bromobenzene	12.982	156	122062	46.847	ug/l	95
78) n-propylbenzene	13.035	91	585903	51.587	ug/l	100
79) 2-Chlorotoluene	13.123	91	362526	48.850	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	424332	52.823	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	55485	47.612	ug/l	92
82) 4-Chlorotoluene	13.217	91	375636	48.097	ug/l	99
83) tert-Butylbenzene	13.435	119	350031	52.649	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	435421	52.518	ug/l	99
85) sec-Butylbenzene	13.612	105	489305	52.102	ug/l	100
86) p-Isopropyltoluene	13.729	119	421834	54.770	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	230686	45.366	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	231568	49.658	ug/l	98
89) n-Butylbenzene	14.053	91	362215	50.276	ug/l	99
90) Hexachloroethane	14.335	117	81515	47.459	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	223772	45.796	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.711	75	29987	45.754	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	119579	44.711	ug/l	99
94) Hexachlorobutadiene	15.500	225	51033	43.451	ug/l	98
95) Naphthalene	15.635	128	384553	48.037	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	116341	44.812	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084571.D
Acq On : 30 Oct 2024 12:09
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/04/2024
Supervised By :Mahesh Dadoda 11/04/2024

Quant Time: Oct 31 18:18:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:18:08 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103024\
Data File : VN084571.D
Acq On : 30 Oct 2024 12:09
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Quant Time: Oct 31 18:18:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:18:08 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/04/2024
Supervised By :Mahesh Dadoda 11/04/2024

