

Surrogate Summary

SDG No.: P4732

Client: Furino and Sons, Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4732-04	PPE-Grab	1,2-Dichloroethane-d4	50	47.6	95	50	163
		Dibromofluoromethane	50	46.9	94	54	147
		Toluene-d8	50	46.7	93	58	134
		4-Bromofluorobenzene	50	41.5	83	29	146
VN1107MBL01	VN1107MBL01	1,2-Dichloroethane-d4	50	48.6	97	50	163
		Dibromofluoromethane	50	50.8	102	54	147
		Toluene-d8	50	46.8	94	58	134
		4-Bromofluorobenzene	50	44.8	90	29	146
VN1107MBS01	VN1107MBS01	1,2-Dichloroethane-d4	50	45.6	91	50	163
		Dibromofluoromethane	50	47.5	95	54	147
		Toluene-d8	50	47.4	95	58	134
		4-Bromofluorobenzene	50	50.4	101	29	146
VN1107MBSD0	VN1107MBSD01	1,2-Dichloroethane-d4	50	45.4	91	50	163
		Dibromofluoromethane	50	47.4	95	54	147
		Toluene-d8	50	47.4	95	58	134
		4-Bromofluorobenzene	50	49.0	98	29	146