

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: FURI01
Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG NO.: P4732
Lab File ID: VN084716.D Date Analyzed: 11/07/2024
Instrument ID: MSVOA_N Time Analyzed: 11:50
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	165434	8.22	276584	9.10	245337	11.87
UPPER LIMIT	330868	8.724	553168	9.6	490674	12.365
LOWER LIMIT	82717	7.724	138292	8.6	122669	11.365
EPA SAMPLE NO.						
PPE-Grab	192832	8.22	338721	9.09	282154	11.87
VN1107MBL01	152647	8.22	260913	9.10	224097	11.87
VN1107MBS01	187169	8.22	306399	9.10	282379	11.87
VN1107MBSD01	175664	8.22	289091	9.10	261498	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.