

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: FURI01
 Lab Code: CHEM Case No.: P4732 SAS No.: P4732 SDG NO.: P4732
 Lab File ID: VN084716.D Date Analyzed: 11/07/2024
 Instrument ID: MSVOA_N Time Analyzed: 11:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	120573	13.788				
UPPER LIMIT	241146	14.288				
LOWER LIMIT	60286.5	13.288				
EPA SAMPLE NO.						
PPE-Grab	119437	13.79				
VN1107MBL01	94404	13.79				
VN1107MBS01	143946	13.79				
VN1107MBSD01	134964	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.