



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms **SDG No.:** P4736
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4736 **SAS No.:** P4736
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	6.16	6.0	103	40 - 160	P	11/08/2024	14:20	LB133365
	Copper	21.4	20.0	107	40 - 160	P	11/08/2024	14:20	LB133365
	Lead	11.8	12.0	98	40 - 160	P	11/08/2024	14:20	LB133365
	Nickel	36.8	40.0	92	40 - 160	P	11/08/2024	14:20	LB133365
	Zinc	41.9	40.0	105	40 - 160	P	11/08/2024	14:20	LB133365
CRA	Mercury	0.21	0.2	103	40 - 160	CV	11/11/2024	14:10	LB133392
CRI01	Cadmium	5.78	6.0	96	40 - 160	P	11/19/2024	11:43	LB133524
	Copper	22.2	20.0	111	40 - 160	P	11/19/2024	11:43	LB133524
	Lead	12.1	12.0	101	40 - 160	P	11/19/2024	11:43	LB133524
	Nickel	39.1	40.0	98	40 - 160	P	11/19/2024	11:43	LB133524
	Zinc	42.4	40.0	106	40 - 160	P	11/19/2024	11:43	LB133524