

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Aramark Uniforms **SDG No.:** P4736
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4736 **SAS No.:** P4736
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	6.11	1.0	611	-5	7	11/08/2024	14:24	LB133365
	Copper	14.3	2.0	715	-18	22	11/08/2024	14:24	LB133365
	Lead	9.97			-12	12	11/08/2024	14:24	LB133365
	Nickel	2.63	2.0	132	-38	42	11/08/2024	14:24	LB133365
	Zinc	7.04			-40	40	11/08/2024	14:24	LB133365
ICSAB01	Cadmium	979	972	101	826	1120	11/08/2024	14:30	LB133365
	Copper	499	511	98	434	588	11/08/2024	14:30	LB133365
	Lead	57.3	49.0	117	37	61	11/08/2024	14:30	LB133365
	Nickel	981	954	103	810	1100	11/08/2024	14:30	LB133365
	Zinc	1090	952	114	809	1095	11/08/2024	14:30	LB133365
ICSA01	Cadmium	3.48	1.0	348	-5	7	11/19/2024	11:47	LB133524
	Copper	7.48	2.0	374	-18	22	11/19/2024	11:47	LB133524
	Lead	8.06			-12	12	11/19/2024	11:47	LB133524
	Nickel	2.06	2.0	103	-38	42	11/19/2024	11:47	LB133524
	Zinc	5.70			-40	40	11/19/2024	11:47	LB133524
ICSAB01	Cadmium	832	972	86	826	1120	11/19/2024	11:51	LB133524
	Copper	493	511	96	434	588	11/19/2024	11:51	LB133524
	Lead	57.8	49.0	118	37	61	11/19/2024	11:51	LB133524
	Nickel	1020	954	107	810	1100	11/19/2024	11:51	LB133524
	Zinc	1090	952	114	809	1095	11/19/2024	11:51	LB133524