

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: P4736

Contract: ARAM01

Lab Code: CHEM

Case No.: P4736

SAS No.: P4736

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164759BS							
Cadmium	ug/L	100	97.8		98	80 - 120	P
Copper	ug/L	150	155		103	80 - 120	P
Lead	ug/L	500	489		98	80 - 120	P
Nickel	ug/L	250	248		99	80 - 120	P
Zinc	ug/L	100	103		103	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: P4736

Contract: ARAM01

Lab Code: CHEM

Case No.: P4736

SAS No.: P4736

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164872BS Mercury	ug/L	4.0	3.62		90	80 - 120	CV