
From: Bett, Daisy <Daisy.Bett@gdit.com>
Sent: Tuesday, November 12, 2024 12:22
To: Deepak Parmar; Sohil Jodhani; Mohammad Ahmed
Cc: burman.jarmael@epa.gov; roberson.sharon@epa.gov; Bauer, Heather E; Johnson, Matthew
Subject: Region 03 | Case 51802 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC | FINAL

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Updated Record of Communication

This ROC is being updated to change the SDG Number referenced in the resolution from C0CE5 to C0CC2.

Good afternoon,


Issue: Regarding SDG C0CC2, Laboratory QC is scheduled for PEST and ARO analyses, but no water samples were designated for QC analysis on the COC and no extra volume was received. The laboratory would like to proceed without Laboratory QC to keep their holding time. Please note that this Case is complete.

Resolution: Per Region 3, the laboratory should note the issue in the SDG Narrative and proceed with the analysis of the samples without performing Laboratory QC for PEST and ARO analyses in SDG C0CC2.

Please note that the laboratory may contact the appropriate CLP PM should any defects need to be waived for this issue.

Thank you,
Daisy Bett
Research Analyst Associate
GDIT Federal Civilian Division
EPA Region 2&3 CLP QSS Coordinator
Under contract to the EPA

T: 571.454.0186
daisy.bett@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



Leave alert: Nov 29th

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Deepak Parmar <Deepak.Parmar@alliancetg.com>
Sent: Tuesday, November 12, 2024 11:26 AM
To: Bett, Daisy <Daisy.Bett@gdit.com>; Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>
Cc: burman.jarmael@epa.gov; roberson.sharon@epa.gov; Bauer, Heather E <Heather.Bauer@gdit.com>; Johnson, Matthew <Matthew.Johnson32@gdit.com>
Subject: RE: Region 03 | Case 51802 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC | FINAL

This Message Is From an External Sender


Please use caution with links, attachments, and any requests for credentials.

Good morning,

Please updated the SDG id C0CE5 to C0CC2.

Thanks & Regards,



Deepak Parmar
QA/QC
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3154
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com     

From: Bett, Daisy <Daisy.Bett@gdit.com>
Sent: Tuesday, November 12, 2024 11:20 AM
To: Deepak Parmar <Deepak.Parmar@AllianceTG.com>; Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>; Mohammad Ahmed <Mohammad.Ahmed@AllianceTG.com>
Cc: burman.jarmael@epa.gov; roberson.sharon@epa.gov; Bauer, Heather E <Heather.Bauer@gdit.com>; Johnson, Matthew <Matthew.Johnson32@gdit.com>
Subject: Region 03 | Case 51802 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC | FINAL

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning,


Issue: Regarding SDG C0CE5, Laboratory QC is scheduled for PEST and ARO analyses, but no water samples were designated for QC analysis on the COC and no extra volume was received. The laboratory would like to proceed without Laboratory QC to keep their holding time. Please note that this Case is complete.

Resolution: Per Region 3, the laboratory should note the issue in the SDG Narrative and proceed with the analysis of the samples without performing Laboratory QC for PEST and ARO analyses in SDG C0CE5.

Please note that the laboratory may contact the appropriate CLP PM should any defects need to be waived for this issue.

Thank you,
Daisy Bett
Research Analyst Associate
GDIT Federal Civilian Division
EPA Region 2&3 CLP QSS Coordinator
Under contract to the EPA

T: 571.454.0186
daisy.bett@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



GENERAL DYNAMICS
Information Technology

Leave alert: Nov 29th

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Burman, Jarmael <Burman.Jarmael@epa.gov>

Sent: Tuesday, November 12, 2024 8:28 AM

To: Bett, Daisy <Daisy.Bett@gdit.com>

Cc: Roberson, Sharon <Roberson.Sharon@epa.gov>

Subject: RE: NEW ISSUE | Case 51802 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning Daisy,

Have ACE make note of the issue in their SDG Narrative and proceed with the analysis of the samples without performing Laboratory QC for Pesticides and Aroclors in SDG C0CE5.

Jarmael Burman
US EPA Region 3 - CLP RR/RSCC/DDS/QA Chemist/DAS PO/EEOC
701 Mapes Road
Fort Meade, Maryland 20755-5350
(410) 305-2743 (office)
(410) 305-3095 (fax)

From: Bett, Daisy <Daisy.Bett@gdit.com>
Sent: Monday, November 11, 2024 2:32 PM
To: Burman, Jarmael <Burman.Jarmael@epa.gov>
Cc: Roberson, Sharon <Roberson.Sharon@epa.gov>
Subject: NEW ISSUE | Case 51802 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.


Good afternoon Jay,

Please see the below issue from ACE.

Issue: Regarding SDG C0CE5, Laboratory QC is scheduled for PEST and ARO analyses, but no water samples were designated for QC on the COC and no extra volume was received. The laboratory would like to proceed without Laboratory QC to keep their holding time. Please note that this Case is complete.

Thank you,
Daisy Bett
Research Analyst Associate
GDIT Federal Civilian Division
EPA Region 2&3 CLP QSS Coordinator
Under contract to the EPA

T: 571.454.0186
daisy.bett@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



GENERAL DYNAMICS
Information Technology

Leave alert: Nov 29th

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Deepak Parmar <Deepak.Parmar@alliancetg.com>
Sent: Saturday, November 9, 2024 12:20 PM
To: Bett, Daisy <Daisy.Bett@gdit.com>
Cc: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Subject: Region 03 | Case 51802 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning,

Lab has open SDG C0CE5 for water samples without lab qc for PEST and PCB analysis , however, a sample was not designated or no extra volume was received. there for The laboratory would like to proceed without Laboratory QC .to keep holding time. Case is complete.

Please see attachment for your reference

Thanks & Regards,



Deepak Parmar
QA/QC
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3154
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

