

## Pratik Patel

---

**From:** Irving, Madison <Madison.Irving@gdit.com>  
**Sent:** Monday, November 25, 2024 2:51 PM  
**To:** Sohil Jodhani; Mohammad Ahmed; Nimisha Pandya  
**Subject:** REGION 8 | CASE 51870 | LAB ACE | SDG H46Q8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL  
**Attachments:** H46R4.pdf; H46S3.pdf; H46X1.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon ACE,

Regarding the laboratory's reported issue for Case 51870/SDG H46Q8, please see the Region's input below and let me know if you have any questions.

**LAB:** "Lab has received water samples for TVOA analysis. Lab has analyzed samples H46X1, H46S3 & H46R4 in a continues analytical sequence. All three samples are found positive and required dilution to bring target analytes within calibration range as you can see attached form-1 with quant reports. Due to continuous analytical sequence, instrument blank was not analyzed in between the samples therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution in final electronic deliverables.

Please see attached."

**REGION:** "The approach is acceptable."

\*\*\*\*\* Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

Madison Irving  
Associate Environmental Analyst  
GDIT Federal Civilian Division  
Under Contract to EPA  
T: 919-768-4149  
15036 Conference Center Drive  
Chantilly, VA 20151  
[www.gdit.com](http://www.gdit.com)

**GENERAL DYNAMICS**  
Information Technology

**LEAVE ALERT: None**

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly

prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

---

**From:** Goodrich, Donald <Goodrich.Donald@epa.gov>

**Sent:** Monday, November 25, 2024 2:46 PM

**To:** Irving, Madison <Madison.Irving@gdit.com>

**Subject:** RE: REGION 8 | CASE 51870 | LAB ACE | SDG H46Q8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

**This Message Is From an External Sender**

Please use caution with links, attachments, and any requests for credentials.

Hi Madison,

The approach is acceptable.

Thanks,  
Don

*Don Goodrich*  
*EPA Region 8*  
*8LSASD-LSB / 5247*  
*303-905-4024 (cell) / [goodrich.donald@epa.gov](mailto:goodrich.donald@epa.gov)*

---

**From:** Irving, Madison <[Madison.Irving@gdit.com](mailto:Madison.Irving@gdit.com)>

**Sent:** Monday, November 25, 2024 11:38 AM

**To:** Goodrich, Donald <[Goodrich.Donald@epa.gov](mailto:Goodrich.Donald@epa.gov)>

**Subject:** REGION 8 | CASE 51870 | LAB ACE | SDG H46Q8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

**Caution:** This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good afternoon Region 8,

Please advise on the laboratory's reported issue below for Case 51870/SDG H46Q8.

**LAB:** "Lab has received water samples for TVOA analysis. Lab has analyzed samples H46X1, H46S3 & H46R4 in a continues analytical sequence. All three samples are found positive and required dilution to bring target analytes within calibration range as you can see attached form-1 with quant reports. Due to continuous analytical sequence, instrument blank was not analyzed in between the samples therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution in final electronic deliverables.

Please see attached."

Thank you,

Madison Irving  
Associate Environmental Analyst  
GDIT Federal Civilian Division  
Under Contract to EPA  
T: 919-768-4149  
15036 Conference Center Drive  
Chantilly, VA 20151  
[www.gdit.com](http://www.gdit.com)

**GENERAL DYNAMICS**  
Information Technology

**LEAVE ALERT: None**

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

---

**From:** DASsupport  
**Sent:** Monday, November 25, 2024 10:43 AM  
**To:** 'Sohil Jodhani' <[Sohil.Jodhani@alliancetg.com](mailto:Sohil.Jodhani@alliancetg.com)>  
**Cc:** Mohammad Ahmed <[mohammad.ahmed@alliancetg.com](mailto:mohammad.ahmed@alliancetg.com)>; Nimisha Pandya <[Nimisha.Pandya@AllianceTG.com](mailto:Nimisha.Pandya@AllianceTG.com)>  
**Subject:** REGION 8 | CASE 51870 | LAB ACE | SDG H46Q8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team  
[DASsupport@gdit.com](mailto:DASsupport@gdit.com)

GDIT Federal Civilian Division  
Under contract to the EPA  
15036 Conference Center Drive  
Chantilly, VA 20151  
[www.gdit.com](http://www.gdit.com)

**GENERAL DYNAMICS**  
Information Technology

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

---

**From:** Sohil Jodhani <[Sohil.Jodhani@alliancetg.com](mailto:Sohil.Jodhani@alliancetg.com)>

**Sent:** Monday, November 25, 2024 10:03 AM

**To:** DASsupport <[DASsupport@gdit.com](mailto:DASsupport@gdit.com)>

**Cc:** Mohammad Ahmed <[mohammad.ahmed@alliancetg.com](mailto:mohammad.ahmed@alliancetg.com)>; Nimisha Pandya <[Nimisha.Pandya@AllianceTG.com](mailto:Nimisha.Pandya@AllianceTG.com)>

**Subject:** REGION 8 | CASE 51870 | LAB ACE | SDG H46Q8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

**This Message Is From an External Sender**

Please use caution with links, attachments, and any requests for credentials.

Hello,

Lab is sending this email with regards to case 51840 and SDG H46Q8.

Lab has received water samples for TVOA analysis. Lab has analyzed samples H46X1, H46S3 & H46R4 in a continues analytical sequence. All three samples are found positive and required dilution to bring target analytes within calibration range as you can see attached form-1 with quant reports. Due to continuous analytical sequence, instrument blank was not analyzed in between the samples therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution in final electronic deliverables.

Please see attached.

**Thanks & Regards,**



**Sohil Jodhani**

**QA/QC Director**

**An Alliance Technical Group Company**

**Main:** 908-789-8900

**Direct:** 908-728-3152

**Address:** 284 Sheffield St, Ste 1, Mountainside, NJ 07092

[www.alliancetg.com](http://www.alliancetg.com)

