

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51863 MA No.: SDG No.: MC0D32
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
		Cr	Cu	Mn	Mo	Ni
Lead	220.353	0	0.0014	0.00034	-0.00148	0.00063

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51863 MA No.: SDG No.: MC0D32
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For			
		Pb	Ti	V	Zn
Lead	220.353	0	0	0	0

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51863 MA No.: SDG No.: MC0D32
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
Lead	220.353	0	0	0	0	