

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51863 MA No. : SDG No.: MC0D32
Initial Calibration Verification Source : MP83143
Continuing Calibration Verification Source : MP83146
Run Batch: LB133447 Analytical Method: ICP-AES
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV011				ID: CCV079				ID: CCV080		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Lead	1000	990	99	0	25000	25000	98	0	24000	98	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51863 MA No. : SDG No.: MC0D32
Initial Calibration Verification Source : MP83143
Continuing Calibration Verification Source : MP83146
Run Batch: LB133447 Analytical Method: ICP-AES
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV081				ID: CCV082		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Lead					25000	25000	99	0	25000	99	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51863 MA No. : SDG No.: MC0D32
Initial Calibration Verification Source : MP83143
Continuing Calibration Verification Source : MP83146
Run Batch: LB133447 Analytical Method: ICP-AES
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV083				ID: CCV084		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Lead					25000	25000	99	1	25000	100	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51863 MA No. : SDG No.: MC0D32
Initial Calibration Verification Source : MP83143
Continuing Calibration Verification Source : MP83146
Run Batch: LB133447 Analytical Method: ICP-AES
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV085				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Lead					25000	25000	101	0			