

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51863 MA No.: SDG No.: MC0D37
Instrument ID: P5 Start Date: 11/18/2024
Analytical Method: ICP-AES Run Batch: LB133486
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Cadmium	0	-0	0	5	5.5	10	625	630	1
Lead	0	0.0	0	10	7.8	-22	6250	6200	-1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51863 MA No.: SDG No.: MC0D37
Instrument ID: P5 Start Date: 11/18/2024
Analytical Method: ICP-AES Run Batch: LB133486
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Cadmium	1250	1200	-4	2500	2500	0	5000	5000	0
Lead	12500	12000	-4	25000	25000	0	50000	50000	0