MC0DA9S

FORM 5A - IN MATRIX SPIKE SAMPLE RECOVERY

Lab Name:	Chemtech Consulting Group			Contract:			
Lab Code:	ACE	Case No.:	51863	MA No. :		SDG No.:	MC0D37
Matrix :	Water			Analytical	Method:	ICP-AES	
% Solids:							

Concentration Units (µg/L, mg/L or mg/kg dry weight): _____ug/L _____

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	۶R	Q
Cadmium	75 - 125	60		1.8	J	50.0	116	
Lead		1100		990		20.0	512	