EPA SAMPLE NO.

MC0DA9L

FORM 8-IN

ICP-AES AND ICP-MS SERIAL DILUTIONS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51863 MA No.: SDG No.: MC0D37

Matrix : Water Analytical Method: ICP-AES

% Solids:

Concentration Units (µg/L, mg/L or mg/kg dry weight): ug/L

Analyte	Initial Sample Result (I)	Q	Serial Dilution Result(S)	Q	% Difference	Q
Cadmium	1.8	J	1.7	J	3	
Lead	990		920		7	