



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc

SDG No.: P4757

Contract: TULL01

Lab Code: CHEM

Case No.: P4757

SAS No.: P4757

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	21.4	20.0	107	40 - 160	P	11/08/2024	14:20	LB133365
	Iron	92.9	100	93	40 - 160	P	11/08/2024	14:20	LB133365
	Lead	11.8	12.0	98	40 - 160	P	11/08/2024	14:20	LB133365