

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>P4757</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>P4757</u> sas no.: <u>P4757</u>
matrix: <u>Water</u>	sample id: <u>P4762-02</u>	client id: <u>EFF-WWMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P4762-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	332		183		150	99		P
Iron	ug/L	75 - 125	1740		115		1500	108		P
Lead	ug/L	75 - 125	515		9.99		500	101		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>P4757</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>P4757</u> sas no.: <u>P4757</u>
matrix: <u>Water</u>	sample id: <u>P4762-02</u>	client id: <u>EFF-WWMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P4762-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	336		183		150	101		P
Iron	ug/L	75 - 125	1780		115		1500	111		P
Lead	ug/L	75 - 125	520		9.99		500	102		P