



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

VERINA CONSULTING GROUP, LLC

Project Name: Rotor Clip - PO# 5183.0001

Project # N/A

Chemtech Project # P4763

Test Name: Residual Chlorine

A. Number of Samples and Date of Receipt:

1 Water sample was received on 11/07/2024.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Ammonia, Metals Group4 and Residual Chlorine. This data package contains results for Residual Chlorine.

C. Analytical Techniques:

The analysis of Residual Chlorine was based on method SM4500 Cl G.

D. QA/ QC Samples:

The Holding Times were met for all samples except for Water Treatment Discharge of Residual Chlorine as samples was receive out of holding time.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (Water Treatment DischargeMS) analysis met criteria for all samples except for Ammonia as N due to Matrix interference.

The Matrix Spike Duplicate (Water Treatment DischargeMSD) analysis met criteria for all samples except for Ammonia as N due to Matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____