



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

VERINA CONSULTING GROUP, LLC

Project Name: Rotor Clip - PO# 5183.0001

Project # N/A

Chemtech Project # P4763

Test Name: Metals Group4

A. Number of Samples and Date of Receipt:

1 Water sample was received on 11/07/2024.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Ammonia, Metals Group4 and Residual Chlorine. This data package contains results for Metals Group4.

C. Analytical Techniques:

The analysis of Metals Group4 was based on method 6010D and digestion based on method 3010 (waters).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate (Water Treatment DischargeMSD) analysis met criteria for all samples except for Zinc due to Chemical interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____