

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	P4763
Project:	Rotor Clip - PO# 5183.0001	Sample ID:	P4763-01
Client ID:	Water Treatment DischargeMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Residual Chlorine	mg/L	71-148	0.43		0.024	J	0.4	1	101		11/08/2024
Ammonia as N	mg/L	75-125	1.00		0.42		1	1	58	*	11/08/2024

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	P4763
Project:	Rotor Clip - PO# 5183.0001	Sample ID:	P4763-01
Client ID:	Water Treatment DischargeMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Residual Chlorine	mg/L	71-148	0.44		0.024	J	0.4	1	104		11/08/2024
Ammonia as N	mg/L	75-125	0.98		0.42		1	1	56	*	11/08/2024