

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	P4763
Project:	Rotor Clip - PO# 5183.0001	Sample ID:	P4763-01
Client ID:	Water Treatment DischargeDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	0.42		0.44		1	5		11/08/2024
Residual Chlorine	mg/L	+/-20	0.024	J	0.026	J	1	8		11/08/2024

Duplicate Sample Summary

Client: VERINA CONSULTING GROUP, LLC	SDG No.: P4763
Project: Rotor Clip - PO# 5183.0001	Sample ID: P4763-01
Client ID: Water Treatment DischargeMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	1.00		0.98		1	2		11/08/2024
Residual Chlorine	mg/L	+/-20	0.43		0.44		1	2.54		11/08/2024