

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : M SC-2

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BLOCK #3 2. N/A

Start Digest Date: 11/08/2024 Time : 12:30 Temp : 96 °C

End Digest Date: 11/08/2024 Time : 15:35 Temp : 96 °C

Digestion tube ID: M6054

Block thermometer ID: MET-DIG. #3

Dig Technician Signature: *JSP*

Supervisor Signature: *[Signature]*

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6000
LFS-2	0.25	M6009
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6116
1:1 HCL	5.00	MP82127
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK #1 Cell #50 : 96 C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/8/24 16:00	<i>JSP Met dig</i>	<i>[Signature]</i>
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
P4763-01	WATER TREATMENT DISCHARGE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	1
P4763-01DUP	WATER TREATMENT DISCHARGE DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	2
P4763-01MS	WATER TREATMENT DISCHARGE MS	<2	50	25	Colorless	Colorless	Clear	Clear	M6000,M6009	3
P4763-01MSD	WATER TREATMENT DISCHARGE MSD	<2	50	25	Colorless	Colorless	Clear	Clear	M6000,M6009	4
PB164796BL	PBW796	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	5
PB164796BS	LCS796	<2	50	25	Colorless	Colorless	Clear	Clear	M6000,M6009	6