

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** P4763
Contract: VERI01 **Lab Code:** CHEM **Case No.:** P4763 **SAS No.:** P4763
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	526	520	101	90 - 110	P	11/11/2024	14:25	LB133395
	Copper	534	510	105	90 - 110	P	11/11/2024	14:25	LB133395
	Nickel	520	530	98	90 - 110	P	11/11/2024	14:25	LB133395
	Zinc	1060	1000	106	90 - 110	P	11/11/2024	14:25	LB133395

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** P4763
Contract: VERI01 **Lab Code:** CHEM **Case No.:** P4763 **SAS No.:** P4763
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	8.58	10.0	86	80 - 120	P	11/11/2024	14:47	LB133395
	Copper	21.4	20.0	107	80 - 120	P	11/11/2024	14:47	LB133395
	Nickel	36.1	40.0	90	80 - 120	P	11/11/2024	14:47	LB133395
	Zinc	42.1	40.0	105	80 - 120	P	11/11/2024	14:47	LB133395

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** P4763
Contract: VERI01 **Lab Code:** CHEM **Case No.:** P4763 **SAS No.:** P4763
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	953	1000	95	90 - 110	P	11/11/2024	15:14	LB133395
	Copper	1250	1250	100	90 - 110	P	11/11/2024	15:14	LB133395
	Nickel	2410	2500	96	90 - 110	P	11/11/2024	15:14	LB133395
	Zinc	2530	2500	101	90 - 110	P	11/11/2024	15:14	LB133395
CCV02	Chromium	931	1000	93	90 - 110	P	11/11/2024	15:37	LB133395
	Copper	1210	1250	97	90 - 110	P	11/11/2024	15:37	LB133395
	Nickel	2340	2500	94	90 - 110	P	11/11/2024	15:37	LB133395
	Zinc	2480	2500	99	90 - 110	P	11/11/2024	15:37	LB133395
CCV03	Chromium	938	1000	94	90 - 110	P	11/11/2024	16:39	LB133395
	Copper	1230	1250	98	90 - 110	P	11/11/2024	16:39	LB133395
	Nickel	2370	2500	95	90 - 110	P	11/11/2024	16:39	LB133395
	Zinc	2490	2500	100	90 - 110	P	11/11/2024	16:39	LB133395
CCV04	Chromium	948	1000	95	90 - 110	P	11/11/2024	17:38	LB133395
	Copper	1280	1250	102	90 - 110	P	11/11/2024	17:38	LB133395
	Nickel	2440	2500	98	90 - 110	P	11/11/2024	17:38	LB133395
	Zinc	2510	2500	101	90 - 110	P	11/11/2024	17:38	LB133395