



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P4763

Contract: VERI01

Lab Code: CHEM

Case No.: P4763

SAS No.: P4763

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Chromium	8.50	10.0	85	40 - 160	P	11/11/2024	14:59	LB133395
	Copper	20.7	20.0	104	40 - 160	P	11/11/2024	14:59	LB133395
	Nickel	36.2	40.0	90	40 - 160	P	11/11/2024	14:59	LB133395
	Zinc	41.1	40.0	103	40 - 160	P	11/11/2024	14:59	LB133395