

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** P4763
Contract: VERI01 **Lab Code:** CHEM **Case No.:** P4763 **SAS No.:** P4763
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	52.2	52.0	100	42	62	11/11/2024	15:04	LB133395
	Copper	15.3	2.0	765	-18	22	11/11/2024	15:04	LB133395
	Nickel	2.14	2.0	107	-38	42	11/11/2024	15:04	LB133395
	Zinc	-1.46			-40	40	11/11/2024	15:04	LB133395
ICSAB01	Chromium	541	542	100	460	624	11/11/2024	15:08	LB133395
	Copper	503	511	98	434	588	11/11/2024	15:08	LB133395
	Nickel	988	954	104	810	1100	11/11/2024	15:08	LB133395
	Zinc	834	952	88	809	1095	11/11/2024	15:08	LB133395