

metals
- 5a -
MATRIX SPIKE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** P4763
contract: VERI01 **lab code:** CHEM **case no.:** P4763 **sas no.:** P4763
matrix: Water **sample id:** P4763-01 **client id:** Water Treatment DischargeMS
Percent Solids for Sample: NA **Spiked ID:** P4763-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	228		47.6		200	90		P
Copper	ug/L	75 - 125	152		14.7		150	91		P
Nickel	ug/L	75 - 125	265		18.4	J	250	99		P
Zinc	ug/L	75 - 125	208		88.2		100	120		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** P4763
contract: VERI01 **lab code:** CHEM **case no.:** P4763 **sas no.:** P4763
matrix: Water **sample id:** P4763-01 **client id:** Water Treatment DischargeMSD
Percent Solids for Sample: NA **Spiked ID:** P4763-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	230		47.6		200	91		P
Copper	ug/L	75 - 125	154		14.7		150	93		P
Nickel	ug/L	75 - 125	268		18.4	J	250	100		P
Zinc	ug/L	75 - 125	216		88.2		100	128	N	P