

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P4763

Contract: VERI01

Lab Code: CHEM

Case No.: P4763 **SAS No.:** P4763

Matrix: Water

Level: LOW

Client ID: Water Treatment DischargeA

Sample ID: P4763-01

Spiked ID: P4763-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Zinc	ug/L	75 - 125	214		88.2		100	126		P