

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** P4763
Contract: VERI01 **Lab Code:** CHEM **Case No.:** P4763 **SAS No.:** P4763
Matrix: Water **Sample ID:** P4763-01 **Client ID:** Water Treatment DischargeDUP
Percent Solids for Sample: NA **Duplicate ID** P4763-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	47.6		50.1		5		P
Copper	ug/L	20	14.7		15.6		6		P
Nickel	ug/L	20	18.4	J	19.6	J	6		P
Zinc	ug/L	20	88.2		95.3		8		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** P4763
Contract: VERI01 **Lab Code:** CHEM **Case No.:** P4763 **SAS No.:** P4763
Matrix: Water **Sample ID:** P4763-01MS **Client ID:** Water Treatment DischargeMSD
Percent Solids for Sample: NA **Duplicate ID** P4763-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	228		230		1		P
Copper	ug/L	20	152		154		1		P
Nickel	ug/L	20	265		268		1		P
Zinc	ug/L	20	208		216		4		P